WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 24 SEP 2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1308 hrs.	1423 hrs.
Sampling Completed (time)	1324 hrs.	1439 hrs.
Sampling Flow Rate	2.5	2.5

Sample Analyses

RemGC1 - 5PPM Cal_ 258-261. CHR (0.9581)
262-265. CHR (0.9680)

Time of GC Calibration: 1252 hrs.

Time Sample Analysis Started: 1308 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1308 hrs.	0.00
	Run 2	1312 hrs.	0.00
	Run 3	1316 hrs.	0.00
	Run 4	1320 hrs.	0.00
	Run 5	1324 hrs.	0.00
Sample B	Run 1	1423 hrs.	0.00
	Run 2	1427 hrs.	0.00
	Run 3	1431 hrs.	0.00
	Run 4	1435 hrs.	0.00
	Run 5	1439 hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) [signature] 3906 (Signature) 24 SEP 2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 01 OCT 2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1405 hrs.	1433 hrs.
Sampling Completed (time)	1421 hrs.	1449 hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5PPM Cal _ 266-269. CHR (0.9887)
270 -273. CHR (0.9794)

Sample Analyses

Time of GC Calibration: 1348 hrs.

Time Sample Analysis Started: _____

		Time of Analyses	Results (ppm)
Sample A	Run 1	1405	0.00
	Run 2	1409	0.00
	Run 3	1413	0.00
	Run 4	1417	0.00
	Run 5	1421	0.00
Sample B	Run 1	1433	0.00
	Run 2	1437	0.00
	Run 3	1441	0.00
	Run 4	1445	0.00
	Run 5	1449	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) _____ 3906 (Signature) 01 OCT 2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 08 OCT2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1352 hrs.	1429 hrs.
Sampling Completed (time)	1408 hrs.	1445 hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1- 5PPM Cal_ 274-277.CHR (0.9743)
278-281.CHR (0.9845)

Sample Analyses

Time of GC Calibration: 1333 hrs

Time Sample Analysis Started: 1352hrs./1429 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1352 hrs.	0.00
	Run 2	1356 hrs.	0.00
	Run 3	1400 hrs.	0.00
	Run 4	1404 hrs.	0.00
	Run 5	1408 hrs.	0.00
Sample B	Run 1	1429 hrs.	0.00
	Run 2	1433 hrs.	0.00
	Run 3	1437 hrs.	0.00
	Run 4	1441 hrs.	0.00
	Run 5	1445 hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 08OCT2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 15 OCT 2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1324hrs.	1345hrs.
Sampling Completed (time)	1340hrs.	1401hrs.
Sampling Flow Rate	2.5	2.5

Sample Analyses

Rem GC1 - 5ppm Cal _ 282-285.CHR (0.9535)
286-289.CHR (0.9635)

Time of GC Calibration: 1257hrs.

Time Sample Analysis Started: 1324hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1324hrs.	0.00
	Run 2	1328hrs.	0.00
	Run 3	1332hrs.	0.00
	Run 4	1336hrs.	0.00
	Run 5	1340hrs.	0.00
Sample B	Run 1	1345hrs.	0.00
	Run 2	1349hrs.	0.00
	Run 3	1353hrs.	0.00
	Run 4	1357hrs.	0.00
	Run 5	1401hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 15OCT2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 22 OCT 2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1212hrs.	1235hrs.
Sampling Completed (time)	1228 hrs.	1251hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 5ppm Cal _ 292-295.CHR (0.9414)
296-299.CHR (0.9581)

Sample Analyses

Time of GC Calibration: 1147hrs.

Time Sample Analysis Started: 1212 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1212hrs.	0.00
	Run 2	1216hrs.	0.00
	Run 3	1220hrs.	0.00
	Run 4	1224hrs.	0.00
	Run 5	1228hrs.	0.00
Sample B	Run 1	1235hrs.	0.00
	Run 2	1239hrs.	0.00
	Run 3	1243hrs.	0.00
	Run 4	1247hrs.	0.00
	Run 5	1251hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 22OCT2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 29 OCT 2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1252hrs.	1320hrs.
Sampling Completed (time)	1308hrs.	1336hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5ppm Cal _ 300 - 303. CHR (0.9570)
304 - 307. CHR (0.9466)

Sample Analyses

Time of GC Calibration: 1220hrs.

Time Sample Analysis Started: 1252hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1252 hrs.	0.00
	Run 2	1256hrs.	0.00
	Run 3	1300hrs.	0.00
	Run 4	1304hrs.	0.00
	Run 5	1308hrs.	0.00
Sample B	Run 1	1320hrs.	0.00
	Run 2	1324hrs.	0.00
	Run 3	1328hrs.	0.00
	Run 4	1332hrs.	0.00
	Run 5	1336hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 29 OCT 2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 05NOV2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1133hrs.	1217hrs.
Sampling Completed (time)	1149hrs.	1233hrs.
Sampling Flow Rate	2.5	2.5

Rem GC2- 5ppm Cal- 308-311.CHR (0.9561)
312-315.CHR (1.001)

Sample Analyses

Time of GC Calibration: 1025hrs

Time Sample Analysis Started: 1133hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1133hrs.	0.00
	Run 2	1137hrs.	0.00
	Run 3	1141hrs.	0.00
	Run 4	1145hrs.	0.00
	Run 5	1149hrs.	0.00
Sample B	Run 1	1217hrs.	0.00
	Run 2	1221hrs.	0.00
	Run 3	1225hrs.	0.00
	Run 4	1229hrs.	0.00
	Run 5	1233hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) [signature] (Signature) 3906 05NOV2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 12NOV2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1324 hrs.	1347 hrs.
Sampling Completed (time)	1340 hrs.	1403 hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5ppm Cal – 316-319. CHR (0.9841)
320-323. CHR (0.9741)

Sample Analyses

Time of GC Calibration: 1305 hrs.

Time Sample Analysis Started: 1324 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1324 hrs.	0.00
	Run 2	1328 hrs.	0.00
	Run 3	1332 hrs.	0.00
	Run 4	1336 hrs.	0.00
	Run 5	1340 hrs.	0.00
Sample B	Run 1	1347 hrs.	0.00
	Run 2	1351 hrs.	0.00
	Run 3	1355 hrs.	0.00
	Run 4	1359 hrs.	0.00
	Run 5	1403 hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 12NOV2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 19NOV2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1451hrs.	1517hrs.
Sampling Completed (time)	1507hrs.	1533hrs.
Sampling Flow Rate	2.5	2.5

RemGC1 - 5ppm Cal _ 324 - 327 . CHR (0.9930)
328 - 331 . CHR (0.9678)

Sample Analyses

Time of GC Calibration: 1404hrs.

Time Sample Analysis Started: 1451hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1451hrs.	0.00
	Run 2	1455hrs.	0.00
	Run 3	1459hrs.	0.00
	Run 4	1503hrs.	0.00
	Run 5	1507hrs.	0.00
Sample B	Run 1	1517hrs.	0.00
	Run 2	1521hrs.	0.00
	Run 3	1525hrs.	0.00
	Run 4	1529hrs.	0.00
	Run 5	1533hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) _____ (Signature) 3906 19NOV2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 03DEC2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1337hrs.	1358hrs.
Sampling Completed (time)	1353hrs.	1414hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5ppm Cal — 332-335.CHR (0.9804)
336-339.CHR (0.9813)

Sample Analyses

Time of GC Calibration: 1312hrs.

Time Sample Analysis Started: 1337hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1337hrs.	0.0
	Run 2	1341hrs.	0.0
	Run 3	1345hrs.	0.0
	Run 4	1349hrs.	0.0
	Run 5	1353hrs.	0.0
Sample B	Run 1	1358hrs.	0.0
	Run 2	1402hrs.	0.0
	Run 3	1406hrs.	0.0
	Run 4	1410hrs.	0.0
	Run 5	1414 hrs.	0.0

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) _____ (Signature) 3906 03DEC2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 10DEC2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0947hrs.	1043hrs.
Sampling Completed (time)	1003hrs.	1059hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5ppm Cal _ 340-343.CHR (0.9803)
344-347.CHR (0.9993)

Sample Analyses

Time of GC Calibration: 0930hrs.

Time Sample Analysis Started: _____

		Time of Analyses	Results (ppm)
Sample A	Run 1	0947hrs.	0.00
	Run 2	0951hrs.	0.00
	Run 3	0955hrs.	0.00
	Run 4	0959hrs.	0.00
	Run 5	1003hrs.	0.00
Sample B	Run 1	1043hrs.	0.00
	Run 2	1047hrs.	0.00
	Run 3	1051hrs.	0.00
	Run 4	1055hrs.	0.00
	Run 5	1059hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) _____ 3906 (Signature) 10DEC2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 17 DEC 2020

	Sample A	Sample B
Ambient Temperature	73.°	73.°
Sampling Initiated (time)	1154 hrs.	1300 hrs.
Sampling Completed (time)	1210 hrs.	1316 hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1-55ppm Cal – 348-351.CHR (0.9685)
352-355.CHR (1.0003)

Sample Analyses

Time of GC Calibration: 1051 hrs.

Time Sample Analysis Started: 1154 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1154 hrs.	0.00
	Run 2	1158 hrs.	0.00
	Run 3	1202 hrs.	0.00
	Run 4	1206 hrs.	0.00
	Run 5	1210 hrs.	0.00
Sample B	Run 1	1300 hrs.	0.00
	Run 2	1304 hrs.	0.00
	Run 3	1308 hrs.	0.00
	Run 4	1312 hrs.	0.00
	Run 5	1316 hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 17 DEC 2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 31 DEC 2020

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1350hrs.	1411hrs.
Sampling Completed (time)	1406hrs.	1427hrs.
Sampling Flow Rate	2.5	2.5

RemGC1- 5ppm Cal. 356-359.CHR (0.9856)
360.363.CHR (0.9956)

Sample Analyses

Time of GC Calibration: 1257hrs.

Time Sample Analysis Started: 1350hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1350hrs.	0.00
	Run 2	1354hrs.	0.00
	Run 3	1358hrs.	0.00
	Run 4	1402hrs.	0.00
	Run 5	1406hrs.	0.00
Sample B	Run 1	1411hrs.	0.00
	Run 2	1415hrs.	0.00
	Run 3	1419hrs.	0.00
	Run 4	1423hrs.	0.00
	Run 5	1427hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) [signature] 3906 (Signature) 31 DEC 2020 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 07JAN 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1009hrs.	1108hrs.
Sampling Completed (time)	1025hrs.	1124hrs.
Sampling Flow Rate	2.5	2.5

Sample Analyses

Rem GC1 - 5ppm Cal — 364-367. CHR (0.975)
369-372. CHR (0.9555)

Time of GC Calibration: 0928hrs.

Time Sample Analysis Started: 1009hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1009hrs.	0.00
	Run 2	1013hrs.	0.00
	Run 3	1017hrs.	0.00
	Run 4	1021hrs.	0.00
	Run 5	1025hrs.	0.00
Sample B	Run 1	1108hrs.	0.00
	Run 2	1112hrs.	0.00
	Run 3	1116hrs.	0.00
	Run 4	1120hrs.	0.00
	Run 5	1124hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 07JAN2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 14JAN2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0432	0510
Sampling Completed (time)	0448	0526
Sampling Flow Rate	2.5	2.5

Rem GC1- 50ppm Cal _ 368-375.CHR ~~to.9458~~ 0.9458)
376.379.CHR 3906 14JAN2021 (0.9814)

Sample Analyses

Time of GC Calibration: 0330hrs.

Time Sample Analysis Started: 0432hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0432	0.00
	Run 2	0436	0.00
	Run 3	0440	0.00
	Run 4	0444	0.00
	Run 5	0448	0.00
Sample B	Run 1	0510	0.00
	Run 2	0514	0.00
	Run 3	0518	0.00
	Run 4	0522	0.00
	Run 5	0526	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 14JAN2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 21 JAN 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0941hrs.	1004hrs.
Sampling Completed (time)	0957hrs.	1020hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5ppm Cal - 385-388.CHR (0.9500)
389-392.CHR (0.9820)

Sample Analyses

Time of GC Calibration: 0901 hrs.

Time Sample Analysis Started: 0941hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0941hrs.	0.00
	Run 2	0945hrs.	0.00
	Run 3	0949hrs.	0.00
	Run 4	0953hrs.	0.00
	Run 5	0957hrs.	0.00
Sample B	Run 1	1004hrs.	0.00
	Run 2	1008hrs.	0.00
	Run 3	1012hrs.	0.00
	Run 4	1016hrs.	0.00
	Run 5	1020hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 21JAN2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 28 JAN 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0816 hrs.	0900
Sampling Completed (time)	0832 hrs.	0916
Sampling Flow Rate	2.5	2.5

RemGC-1 5ppm Cal — 394 - 397. CHR (0.9621)
399 - 402. CHR (0.9372)

Sample Analyses

Time of GC Calibration: 0724 hrs.

Time Sample Analysis Started: 0816 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0816 hrs.	0.00
	Run 2	0820 hrs.	0.00
	Run 3	0824 hrs.	0.00
	Run 4	0828 hrs.	0.00
	Run 5	0832 hrs.	0.00
Sample B	Run 1	0900 hrs.	0.00
	Run 2	0904 hrs.	0.00
	Run 3	0908 hrs.	0.00
	Run 4	0912 hrs.	0.00
	Run 5	0916 hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 28 JAN 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 04FEB2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1152hrs.	1303hrs.
Sampling Completed (time)	1208hrs.	1319hrs.
Sampling Flow Rate	2.5	2.5

Rem GC-1 5ppm Cal_404.407.CHR(0.9779)
409.412.CHR(0.9559)

Sample Analyses

Time of GC Calibration: 1130hrs.

Time Sample Analysis Started: 1152hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1152hrs.	0.00
	Run 2	1156hrs.	0.00
	Run 3	1200hrs.	0.00
	Run 4	1204hrs.	0.00
	Run 5	1208hrs.	0.00
Sample B	Run 1	1303hrs.	0.00
	Run 2	1307hrs.	0.00
	Run 3	1311hrs.	0.00
	Run 4	1315hrs.	0.00
	Run 5	1319hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) [signature] 3906 (Signature) 04FEB2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 11 FEB 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0956hrs.	1029hrs.
Sampling Completed (time)	1012hrs.	1045hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1-5ppm Cal - 414-417.CHR (0.9435)
420-423.CHR (0.9690)

Sample Analyses

Time of GC Calibration: 0826hrs.

Time Sample Analysis Started: 0956hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0956hrs.	0.00
	Run 2	1000hrs.	0.00
	Run 3	1004hrs.	0.00
	Run 4	1008hrs.	0.00
	Run 5	1012hrs.	0.00
Sample B	Run 1	1029hrs.	0.00
	Run 2	1033hrs.	0.00
	Run 3	1037hrs.	0.00
	Run 4	1041hrs.	0.00
	Run 5	1045hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 11FEB2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 18FEB2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0931hrs.	0953hrs.
Sampling Completed (time)	0947hrs.	1009 hrs.
Sampling Flow Rate	2.5	2.5

Rem GC 1.5ppm Cal _ 424-427.CHR (0.9430)
429-432.CHR (0.9247)

Sample Analyses

Time of GC Calibration: 0835hrs.

Time Sample Analysis Started: 0931 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0931hrs.	0.00
	Run 2	0935hrs.	0.00
	Run 3	0939hrs.	0.00
	Run 4	0943hrs.	0.00
	Run 5	0947hrs.	0.00
Sample B	Run 1	0953hrs.	0.00
	Run 2	0957hrs.	0.00
	Run 3	1001hrs.	0.00
	Run 4	1005 hrs.	0.00
	Run 5	1009hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name)  3906 (Signature) 18FEB2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 25FEB2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1113hrs.	1143hrs.
Sampling Completed (time)	1129hrs.	1159hrs.
Sampling Flow Rate	2.5	2.5

RemGC1- 5ppm Cal - .434 - 437.CHR (0.9661)
439 - 442.CHR (0.9630)

Sample Analyses

Time of GC Calibration: 1022hrs.

Time Sample Analysis Started: 1113hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1113hrs.	0.00
	Run 2	1117hrs.	0.00
	Run 3	1121hrs.	0.00
	Run 4	1125hrs.	0.00
	Run 5	1129hrs.	0.00
Sample B	Run 1	1143hrs.	0.00
	Run 2	1147hrs.	0.00
	Run 3	1151hrs.	0.00
	Run 4	1155hrs.	0.00
	Run 5	1159hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 25FEB2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 04MAR2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1013hrs.	1035hrs.
Sampling Completed (time)	1029hrs.	1051hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1. 5ppm Cal — 444-447. CHR (0.9487)
449-452. CHR (0.9284)

Sample Analyses

Time of GC Calibration: 0931hrs.

Time Sample Analysis Started: 1013hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1013hrs.	0.00
	Run 2	1017hrs.	0.00
	Run 3	1021hrs.	0.00
	Run 4	1025hrs. ~~1025hrs.~~ KS 3906 04MAR2021	0.00
	Run 5	1029hrs.	0.00
Sample B	Run 1	1035hrs.	0.00
	Run 2	1039hrs.	0.00
	Run 3	1043hrs.	0.00
	Run 4	1047hrs.	0.00
	Run 5	1051hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) _____ (Signature) 3906 04MAR2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 11MAR2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0817hrs.	0844hrs.
Sampling Completed (time)	0833hrs.	0900hrs.
Sampling Flow Rate	2.5	2.5

Rem GC 1.5ppm cal — 454-457.CHR (0.9421)
459-462.CHR (0.9246)

Sample Analyses

Time of GC Calibration: 0753hrs.

Time Sample Analysis Started: 0817hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0817hrs.	0.00
	Run 2	0821hrs.	0.00
	Run 3	0825hrs.	0.00
	Run 4	0829hrs.	0.00
	Run 5	0833hrs.	0.00
Sample B	Run 1	0844hrs.	0.00
	Run 2	0848hrs.	0.00
	Run 3	0852hrs.	0.00
	Run 4	0856hrs.	0.00
	Run 5	0900hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 11MAR2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 18 MAR 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1024 hrs.	1054 hrs.
Sampling Completed (time)	1040 hrs.	1110 hrs.
Sampling Flow Rate	2.5	2.5

Sample Analyses

Rem GC-1 - 5ppm Cal — 464-467. CHR (0.9315)
469-472. CHR (0.8876)

Time of GC Calibration: 0801 hrs.

Time Sample Analysis Started: 1024 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1024 hrs.	0.00
	Run 2	1028 hrs.	0.00
	Run 3	1032 hrs.	0.00
	Run 4	1036 hrs.	0.00
	Run 5	1040 hrs.	0.00
Sample B	Run 1	1054 hrs.	0.00
	Run 2	1058 hrs.	0.00
	Run 3	1102 hrs.	0.00
	Run 4	1106 hrs.	0.00
	Run 5	1110 hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 18 MAR 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 25MAR2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0945hrs.	1017hrs.
Sampling Completed (time)	1001hrs.	1033hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5ppm Cal — 474-477. CHR (0.9591)
479-482. CHR (0.9195)

Sample Analyses

Time of GC Calibration: 0920hrs.

Time Sample Analysis Started: 0945hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0945hrs.	0.00
	Run 2	0949hrs.	0.00
	Run 3	0953hrs.	0.00
	Run 4	0957hrs.	0.00
	Run 5	1001hrs.	0.00
Sample B	Run 1	1017hrs.	0.00
	Run 2	1021hrs.	0.00
	Run 3	1025hrs.	0.00
	Run 4	1029hrs.	0.00
	Run 5	1033hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) 3906 (Signature) 25MAR2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 01APR2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1027hrs.	1055hrs.
Sampling Completed (time)	1043hrs.	1111 hrs.
Sampling Flow Rate	2.5	2.5

RemGC1 - 5ppm Cal — 484-487.CHR (0.9515)
489-492.CHR (0.9755)

Sample Analyses

Time of GC Calibration: 0959hrs.

Time Sample Analysis Started: 1027hrs

		Time of Analyses	Results (ppm)
Sample A	Run 1	1027hrs.	0.00
	Run 2	1031hrs.	0.00
	Run 3	1035hrs.	0.00
	Run 4	1039hrs.	0.00
	Run 5	1043hrs.	0.00
Sample B	Run 1	1055hrs.	0.00
	Run 2	1059hrs.	0.00
	Run 3	1103hrs.	0.00
	Run 4	1107hrs.	0.00
	Run 5	1111hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Kevin Smith (Employee Name) [signature] 3906 (Signature) 01APR2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 08APR2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1411hrs.	1454hrs.
Sampling Completed (time)	1427hrs.	1510hrs.
Sampling Flow Rate	2.5	2.5

Sample Analyses

Ran GC1 - 5ppm Cal — 494-497.CHR (0.9538)
499-502.CHR (0.9429)

Time of GC Calibration: 1259hrs.

Time Sample Analysis Started: 1411hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1411 hrs.	0.00
	Run 2	1415hrs.	0.00
	Run 3	1419hrs.	0.00
	Run 4	1423hrs.	0.00
	Run 5	1427hrs.	0.00
Sample B	Run 1	1454hrs.	0.00
	Run 2	1458hrs.	0.00
	Run 3	1502hrs.	0.00
	Run 4	1506 hrs.	0.00
	Run 5	1510 hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name)  3906 (Signature) 08APR2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 15 APR 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0504hrs.	0551hrs.
Sampling Completed (time)	0520hrs.	0607hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 · 5ppm Cal _ 504 - 507.CHR (0.9666)
509 - 512.CHR (0.9413)

Sample Analyses

Time of GC Calibration: 0411hrs.

Time Sample Analysis Started: 0504hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0504 hrs.	0.00
	Run 2	0508hrs.	0.00
	Run 3	0512hrs.	0.00
	Run 4	0516hrs.	0.00
	Run 5	0520hrs.	0.00
Sample B	Run 1	0551hrs.	0.00
	Run 2	0555hrs.	0.00
	Run 3	0559hrs.	0.00
	Run 4	0603hrs.	0.00
	Run 5	0607hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 15 APR 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 23APR2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0940hrs.	1009hrs.
Sampling Completed (time)	0956hrs.	1025hrs.
Sampling Flow Rate	2.5	2.5

Sample Analyses

RemGC1 - 5ppm Cal _ 514-517.CHR (0.9639)
520-523.CHR (0.9624)

Time of GC Calibration: 0905hrs.

Time Sample Analysis Started: 0940hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0940hrs.	0.00
	Run 2	0944hrs.	0.00
	Run 3	0948hrs.	0.00
	Run 4	0952hrs.	0.00
	Run 5	0956hrs.	0.00
Sample B	Run 1	1009hrs.	0.00
	Run 2	1013hrs.	0.00
	Run 3	1017hrs.	0.00
	Run 4	1021hrs.	0.00
	Run 5	1025hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 23APR2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 30APR2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0447hrs.	0515hrs.
Sampling Completed (time)	0503hrs.	0531hrs.
Sampling Flow Rate	2.5	2.5

Sample Analyses

Rem GC1-5ppmCal — 524-527. CHR (0.9867)
529-532. CHR (0.9409)

Time of GC Calibration: 0352hrs.

Time Sample Analysis Started: 0447hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0447hrs.	0.00
	Run 2	0451hrs.	0.00
	Run 3	0455hrs.	0.00
	Run 4	0459hrs.	0.00
	Run 5	0503hrs.	0.00
Sample B	Run 1	0515hrs.	0.00
	Run 2	0519hrs.	0.00
	Run 3	0523hrs.	0.00
	Run 4	0527hrs.	0.00
	Run 5	0531hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 30APR2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 07MAY2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0803hrs.	0825hrs.
Sampling Completed (time)	0819hrs.	0841hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5ppm Cal _ 533 - 536. CHR (0.9710)
538 - 541. CHR (0.9612)

Sample Analyses

Time of GC Calibration: 0713hrs.

Time Sample Analysis Started: 0803hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0803hrs.	0.00
	Run 2	0807hrs.	0.00
	Run 3	0811hrs.	0.00
	Run 4	0815hrs.	0.00
	Run 5	0819hrs.	0.00
Sample B	Run 1	0825hrs.	0.00
	Run 2	0829hrs.	0.00
	Run 3	0833hrs.	0.00
	Run 4	0837hrs.	0.00
	Run 5	0841hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 07MAY2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: ~~14~~ 13MAY2021 KS 3906 13 MAY 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0923hrs.	1041hrs.
Sampling Completed (time)	0939hrs.	1057hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5ppm Cal. — 543-546. CHR (0.9545)
548-551. CH R (0.9354)

Sample Analyses

Time of GC Calibration: 0803hrs.

Time Sample Analysis Started: 0923hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0923hrs.	0.00
	Run 2	0927hrs.	0.00
	Run 3	0931hrs.	0.00
	Run 4	0935hrs.	0.00
	Run 5	0939hrs.	0.00
Sample B	Run 1	1041hrs.	0.00
	Run 2	1045hrs ~~1046hrs.~~	0.00
	Run 3	1049hrs. ~~1050hrs.~~	0.00
	Run 4	1053hrs.	0.00
	Run 5	1057hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 13MAY2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

KS 3906
13 MAY 2021

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: ~~20 MAY 2021~~ 21 MAY 2021 KS 3906 21 MAY 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0726hrs.	0752hrs.
Sampling Completed (time)	0742hrs.	0808hrs.
Sampling Flow Rate	2.5	2.5

RemGC1 - 5ppm Cal _ 558-561. CHR (0.8913)
563-566. CHR (0.9047)

Sample Analyses

Time of GC Calibration: 0707hrs.

Time Sample Analysis Started: 0726hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0726hrs.	0.00
	Run 2	0730hrs.	0.00
	Run 3	0734hrs.	0.00
	Run 4	0738hrs.	0.00
	Run 5	0742hrs.	0.00
Sample B	Run 1	0752hrs.	0.00
	Run 2	0756hrs.	0.00
	Run 3	0800hrs.	0.00
	Run 4	0804hrs.	0.00
	Run 5	0808hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 21 MAY 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 27MAY2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1141hrs.	1209hrs.
Sampling Completed (time)	1157hrs.	1225hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1- 5ppm Cal _ 573-576.CHR (0.9507)
578-581.CHR (0.9567)

Sample Analyses

Time of GC Calibration: 1058hrs.

Time Sample Analysis Started: 1141hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1141hrs.	0.00
	Run 2	1145hrs.	0.00
	Run 3	1149hrs.	0.00
	Run 4	1153hrs.	0.00
	Run 5	1157hrs.	0.00
Sample B	Run 1	1209hrs.	0.00
	Run 2	1213hrs.	0.00
	Run 3	1217hrs.	0.00
	Run 4	1221hrs.	0.00
	Run 5	1225hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 27MAY2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 03June2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	1101 hrs.	1121 hrs.
Sampling Completed (time)	1117hrs.	1137hrs.
Sampling Flow Rate	2.5	2.5

Rem Gc C1- 5ppm Cal — 583 - 586. CHR (0.9242)
588 - 591. CHR (1.0309)

Sample Analyses

Time of GC Calibration: 0803hrs

Time Sample Analysis Started: 1101 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1101hrs.	0.00
	Run 2	1105hrs.	0.00
	Run 3	1109hrs.	0.00
	Run 4	1113hrs.	0.00
	Run 5	1117hrs.	0.00
Sample B	Run 1	1121hrs.	0.00
	Run 2	1125hrs.	0.00
	Run 3	1129hrs.	0.00
	Run 4	1133hrs.	0.00
	Run 5	1137hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 03June2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 11 Jun 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0759hrs.	0909hrs.
Sampling Completed (time)	0815hrs.	0925hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5ppm Cal _ 593-596. CHR (1.161)
598-601. CHR (1.727)

Sample Analyses

Time of GC Calibration: 0720hrs.

Time Sample Analysis Started: 0759hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0759hrs.	0.00
	Run 2	0803hrs.	0.00
	Run 3	0807hrs.	0.00
	Run 4	0811hrs.	0.00
	Run 5	0815hrs.	0.00
Sample B	Run 1	0909hrs.	0.00
	Run 2	0913hrs.	0.00
	Run 3	0917hrs.	0.00
	Run 4	0921hrs.	0.00
	Run 5	0925hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 11 JUN 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 17Jun2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0912hrs.	0945hrs.
Sampling Completed (time)	0928hrs.	1001 hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1- 5ppm Cal _ 603-606. CHR (1.021)
608-611. CHR (1.0033)

Sample Analyses

Time of GC Calibration: 0720hrs.

Time Sample Analysis Started: 0912hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0912hrs.	0.00
	Run 2	0916hrs.	0.00
	Run 3	0920hrs.	0.00
	Run 4	0924hrs.	0.00
	Run 5	0928hrs.	0.00
Sample B	Run 1	0945hrs.	0.00
	Run 2	0949hrs.	0.00
	Run 3	0953hrs.	0.00
	Run 4	0957hrs.	0.00
	Run 5	1001hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) [signature] 3906 (Signature) 17JUN2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 25 Jun 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0432hrs.	0522hrs.
Sampling Completed (time)	0448hrs.	0538hrs.
Sampling Flow Rate	2.5	2.5

Rem GC1 - 5ppm Cal _ 613-616.CHR (1.0278)
618-621.CHR (1.0037)

Sample Analyses

Time of GC Calibration: 0338hrs.

Time Sample Analysis Started: _____

		Time of Analyses	Results (ppm)
Sample A	Run 1	0432hrs.	0.00
	Run 2	0436hrs.	0.00
	Run 3	0440hrs.	0.00
	Run 4	0444hrs.	0.00
	Run 5	0448hrs.	0.00
Sample B	Run 1	0522hrs.	0.00
	Run 2	0526hrs.	0.00
	Run 3	0530hrs.	0.00
	Run 4	0534hrs.	0.00
	Run 5	0538hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 25 Jun 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Brad Stout Initials BS

GC Analyses By: Brad Stout Initials BS

Sample Collection Data

Monitoring Date: 02Jul 2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0811 hrs	0831 hrs
Sampling Completed (time)	0827 hrs	0847 hrs
Sampling Flow Rate	2.5	2.5

Sample Analyses

Rem GC 1-5ppm Cal 628-632.CHR (0.9753)
633-637.CHR (1.0055)

Time of GC Calibration: 0738hrs

Time Sample Analysis Started: 0811 hrs

		Time of Analyses	Results (ppm)
Sample A	Run 1	0811 hrs	0.00
	Run 2	0815hrs	0.00
	Run 3	0819 hrs	0.00
	Run 4	0823 hrs	0.00
	Run 5	0827 hrs	0.00
Sample B	Run 1	0831hrs	0.00
	Run 2	0835hrs	0.00
	Run 3	0839 hrs	0.00
	Run 4	0843 hrs	0.00
	Run 5	0847 hrs	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Brad Stout (Employee Name) Brad Stout CECI (Signature) 02Jul 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 16 Jul 2021

	Sample A	Sample B
Ambient Temperature	73°	73° ~~0945hrs.~~ EC 3906
Sampling Initiated (time)	0744 hrs.	0945 hrs.
Sampling Completed (time)	0800 hrs.	1001 hrs.
Sampling Flow Rate	2.5	2.5

Rem GC 1 · 5ppm Cal _ 638-641. CHR (0.8684)
643-646. CHR (0.9240)

Sample Analyses

Time of GC Calibration: 0715 hrs.

Time Sample Analysis Started: 0744 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0744 hrs.	0.00
	Run 2	0748 hrs.	0.00
	Run 3	0752 hrs.	0.00
	Run 4	0756 hrs.	0.00
	Run 5	0800 hrs.	0.00
Sample B	Run 1	0945 hrs.	0.00
	Run 2	0949 hrs.	0.00
	Run 3	0953 hrs.	0.00
	Run 4	0957 hrs.	0.00
	Run 5	1001 hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 16 Jul 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Kevin Smith Initials KS

GC Analyses By: Kevin Smith Initials KS

Sample Collection Data

Monitoring Date: 23JUL2021

	Sample A	Sample B
Ambient Temperature	73°	73°
Sampling Initiated (time)	0738hrs.	0810hrs.
Sampling Completed (time)	0754hrs.	0826hrs.
Sampling Flow Rate	2.5	2.5

Sample Analyses

Rem GC 1 - 5ppm Cal _ 653-656. CH2 (0.8449)
658-661. CH2 (1.1703)

Time of GC Calibration: 0713hrs.

Time Sample Analysis Started: 0738hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	0738hrs.	0.00
	Run 2	0742hrs.	0.00
	Run 3	0746hrs.	0.00
	Run 4	0750hrs.	0.00
	Run 5	0754hrs.	0.00
Sample B	Run 1	0810hrs.	0.00
	Run 2	0814hrs.	0.00
	Run 3	0818hrs.	0.00
	Run 4	0822hrs.	0.00
	Run 5	0826hrs.	0.00

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: K. Smith (Employee Name) 3906 (Signature) 23JUL2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Daniel Lowers Brad Stout Initials DL BS

GC Analyses By: Daniel Lowers Brad Stout Initials DL BS

Sample Collection Data

Monitoring Date: 29 Jul 2021

	Sample A	Sample B
Ambient Temperature	72°F	72° F
Sampling Initiated (time)	13:40	14:01
Sampling Completed (time)	13:59	14:20
Sampling Flow Rate	2.5 ml/min	2.5 ml/mh

Sample Analyses

Time of GC Calibration: 08:00

Time Sample Analysis Started: 13:40

		Time of Analyses	Results (ppm)
Sample A	Run 1	13:43	0.0000
	Run 2	13:47	0.0000
	Run 3	13:51	0.0000
	Run 4	13:55	0.0000
	Run 5	13:59	0.0000
Sample B	Run 1	14:04	0.0000
	Run 2	14:08	0.0000
	Run 3	14:12	0.0000
	Run 4	14:16	0.0000
	Run 5	14:20	0.0000

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Daniel Lowers (Employee Name) Daniel Lowers (Signature) 30 Jul 2021 (Date)

Notes: Brad Stout Brad Stout 30 Jul 2021

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Daniel Lowers Initials DL

GC Analyses By: Daniel Lowers Initials DL

Sample Collection Data

Monitoring Date: 05 Aug 2021

	Sample A	Sample B
Ambient Temperature	72°F	72°F
Sampling Initiated (time)	12:48	13:09
Sampling Completed (time)	13:07	13:28
Sampling Flow Rate	2.5 mL/min	2.5 mL/min

Sample Analyses

Time of GC Calibration: 9:30

Time Sample Analysis Started: 12:48

		Time of Analyses	Results (ppm)
Sample A	Run 1	12:51	0.0000
	Run 2	12:55	0.0000
	Run 3	12:59	0.0000
	Run 4	13:03	0.0000
	Run 5	13:07	0.0000
Sample B	Run 1	13:12	0.0000
	Run 2	13:16	0.0000
	Run 3	13:20	0.0000
	Run 4	13:24	0.0000
	Run 5	13:28	0.0000

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Daniel Lowers (Employee Name) Daniel Lowers (Signature) 05 Aug 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Daniel Lowers Initials DL

GC Analyses By: Daniel Lowers Initials DL

Sample Collection Data

Monitoring Date: 12 Aug 2021

	Sample A	Sample B
Ambient Temperature	72 °F	72 °F
Sampling Initiated (time)	12:26	12:46
Sampling Completed (time)	12:45	13:01
Sampling Flow Rate	2.5 mL/min	2.5 mL/min

Sample Analyses

Time of GC Calibration: 9:30

Time Sample Analysis Started: 12:26

		Time of Analyses	Results (ppm)
Sample A	Run 1	12:29	0.0000
	Run 2	12:33	0.0000
	Run 3	12:37	0.0000
	Run 4	12:41	0.0000
	Run 5	12:45	0.0000
Sample B	Run 1	12:49	0.0000
	Run 2	12:52	0.0000
	Run 3	12:55	0.0000
	Run 4	12:58	0.0000
	Run 5	13:01	0.0000

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Daniel Lowers (Employee Name) Daniel Lowers (Signature) 12 Aug 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Daniel Lowers Initials DL

GC Analyses By: Daniel Lowers Initials DL

Sample Collection Data

Monitoring Date: 19Aug2021

	Sample A	Sample B
Ambient Temperature	75°F	75°F
Sampling Initiated (time)	11:23	11:41
Sampling Completed (time)	11:38	11:56
Sampling Flow Rate	2.5 mL/min	2.5 mL/min

Sample Analyses

Time of GC Calibration: 9:30

Time Sample Analysis Started: 11:23

		Time of Analyses	Results (ppm)
Sample A	Run 1	11:26	0.0000
	Run 2	11:29	0.0000
	Run 3	11:32	0.0000
	Run 4	11:35	0.0000
	Run 5	11:38	0.0000
Sample B	Run 1	11:44	0.0000
	Run 2	11:47	0.0000
	Run 3	11:50	0.0000
	Run 4	11:53	0.0000
	Run 5	11:56	0.0000

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Daniel Lowers (Employee Name) Daniel Lowers (Signature) 19Aug2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Daniel Lowers Initials DL

GC Analyses By: Daniel Lowers Initials DL

Sample Collection Data

Monitoring Date: 26Aug2021

	Sample A	Sample B
Ambient Temperature	75°F	75°F
Sampling Initiated (time)	14:32	14:52
Sampling Completed (time)	14:47	15:07
Sampling Flow Rate	2.5 mL/min	2.5 mL/min

Sample Analyses

Time of GC Calibration: 10:09

Time Sample Analysis Started: 14:32

		Time of Analyses	Results (ppm)
Sample A	Run 1	14:35	0.0000
	Run 2	14:38	0.0000
	Run 3	14:41	0.0000
	Run 4	14:44	0.0000
	Run 5	14:47	0.0000
Sample B	Run 1	14:55	0.0000
	Run 2	14:58	0.0000
	Run 3	15:01	0.0000
	Run 4	15:04	0.0000
	Run 5	15:07	0.0000

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Daniel Lowers (Employee Name) Daniel Lowers (Signature) 26Aug2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Daniel Lowers Initials DL

GC Analyses By: Daniel Lowers Initials DL

Sample Collection Data

Monitoring Date: 02 Sep 2021

	Sample A	Sample B
Ambient Temperature	75°F	75°F
Sampling Initiated (time)	12:19	12:35
Sampling Completed (time)	12:34	
Sampling Flow Rate	2.5 mL/min	2.5 mL/min

Sample Analyses

Time of GC Calibration: 9:53

Time Sample Analysis Started: 12:19

		Time of Analyses	Results (ppm)
Sample A	Run 1	12:22	0.0000
	Run 2	12:25	0.0000
	Run 3	12:28	0.0000
	Run 4	12:31	0.0000
	Run 5	12:34	0.0000
Sample B	Run 1	12:38	0.0000
	Run 2	12:41	0.0000
	Run 3	12:44	0.0000
	Run 4	12:47	0.0000
	Run 5	12:50	0.0000

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Daniel Lowers (Employee Name) Daniel Lowers (Signature) 02 Sep 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Daniel Lowers Initials DL

GC Analyses By: Daniel Lowers Initials DL

Sample Collection Data

Monitoring Date: 09 Sep 2021

	Sample A	Sample B
Ambient Temperature	75 °F	75 °F
Sampling Initiated (time)	13:55	14:19
Sampling Completed (time)	14:11	14:34
Sampling Flow Rate	2.5 ml/min	2.5 ml/min

Sample Analyses

Time of GC Calibration: 8:06

Time Sample Analysis Started: 13:58

		Time of Analyses	Results (ppm)
Sample A	Run 1	13:58	0.0000
	Run 2	14:01	0.0000
	Run 3	14:04	0.0000
	Run 4	14:08	0.0000
	Run 5	14:11	0.0000
Sample B	Run 1	14:22	0.0000
	Run 2	14:25	0.0000
	Run 3	14:28	0.0000
	Run 4	14:31	0.0000
	Run 5	14:34	0.0000

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Daniel Lowers (Employee Name) Daniel Lowers (Signature) 09 Sep 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Daniel Lowers Initials DL

GC Analyses By: Daniel Lowers Initials DL

Sample Collection Data

Monitoring Date: 16 Sep 2021

	Sample A	Sample B
Ambient Temperature	75 °F	75 °F
Sampling Initiated (time)	11:55	12:11
Sampling Completed (time)	12:10	12:26
Sampling Flow Rate	2.5 mL/min	2.5 mL/min

Sample Analyses

Time of GC Calibration: 8:00

Time Sample Analysis Started: 11:55

		Time of Analyses	Results (ppm)
Sample A	Run 1	11:58	0.0000
	Run 2	12:01	0.0000
	Run 3	12:04	0.0000
	Run 4	12:07	0.0000
	Run 5	12:10	0.0000
Sample B	Run 1	12:14	0.0000
	Run 2	12:17	0.0000
	Run 3	12:20	0.0000
	Run 4	12:23	0.0000
	Run 5	12:26	0.0000

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Daniel Lowers (Employee Name) Daniel Lowers (Signature) 16 Sep 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003

WEEKLY ARV MONITORING LOG
COOK STERILIZATION FACILITY
ELLETTSVILLE, INDIANA

Aeration Room Vent Monitoring

Weekly Log Sheet Parameters & Results

Sample Collection By: Brad Stout Initials BS

GC Analyses By: Brad Stout Initials BS

Sample Collection Data

Monitoring Date: 23 Sep 2021

	Sample A	Sample B
Ambient Temperature	75°F	75°F
Sampling Initiated (time)	1032 hrs	1059 hrs
Sampling Completed (time)	1047 hrs	1114 hrs
Sampling Flow Rate	2.5 mL/mn	2.5 mL/mn

Sample Analyses

Time of GC Calibration: 0657 hrs

Time Sample Analysis Started: 1032 hrs.

		Time of Analyses	Results (ppm)
Sample A	Run 1	1035 hrs	0.0000
	Run 2	1038 hrs	0.0000
	Run 3	1041 hrs	0.0000
	Run 4	1044 hrs	0.0000
	Run 5	1047 hrs	0.0000
Sample B	Run 1	1102 hrs	0.0000
	Run 2	1105 hrs	0.0000
	Run 3	1108 hrs	0.0000
	Run 4	1111 hrs	0.0000
	Run 5	1114 hrs	0.0000

Are at least three analytical results within 5% of the average detection? (Y/N) Y

Cook Certification: Brad Stout (Employee Name) Brad Stout (Signature) 23 Sep 2021 (Date)

Notes:

1) Refer to Performance Specifications and Quality Assurance Procedures within the Recommended Operating Parameters for Ethylene Oxide Sterilization Emission Control Equipment dated May 29, 2003